NOTES:


1. SUBSTRATE: FUSED SILICA


2. RoHS: COMPLIANT

3. SURFACE S2 TO BE PARALLEL TO SURFACE \$1 TO WITHIN < 5 ARCMIN

4. COATING (APPLY ACROSS COATING APERTURE) S1& S2: R(AVG) ≤0.5% FROM 600 - 1050nm







FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

| SHAPE | \$1 PLANO | S2 PLANO | | | | Edmund Optic | CS® |
|------------------|----------------------|----------------------|------------------------|----|--------|--|-----------------|
| SURFACE ACCURACY | 1λ | 1λ | | 1 | | 25 x 25mm NIR I, 1λ FUSED SILICA WINDO | 7 W/ |
| SURFACE QUALITY | 60-40 | 60-40 | THIRD ANGLE PROJECTION | | TITLE | 23 X 23ITIITI NIK I, IX I OSED SIEICA WINDOW | |
| CLEAR APERTURE | 22.50 x 22.50 | 22.50 x 22.50 | | | | | CHEET |
| BEVEL | PROTECTIVE AS NEEDED | PROTECTIVE AS NEEDED | ALL DIMS IN | mm | DWG NO | 36979 | SHEET 1 OF 1 |

