## NOTES:

- SUBSTRATE: FUSED SILICA
- 2. SURFACE S2 TO BE PARALLEL TO S1 TO WITHIN <3 ARCMIN
- 3. COATING (APPLY ACROSS COATING APERTURE)

S1: T(ABS) > 70% @ 266nm @ 45° AOI R(ABS) > 99% @ 532nm @ 45° AOI R(ABS) > 99% @ 1064nm @ 45° AOI


DAMAGE THRESHOLD, PULSED: 7.5 J/cm<sup>2</sup> @ 20ns, 20Hz @ 1064nm 5 J/cm<sup>2</sup> @ 20ns, 20Hz @ 532nm 1 J/cm<sup>2</sup> @ 20ns, 20Hz @ 266nm


S2: R(ABS) < 1.0% @ 266nm @ 45° AOI

DAMAGE THRESHOLD, PULSED: 1 J/cm<sup>2</sup> @ 20ns, 20Hz @ 266nm

FINE GRIND SURFACE

5. ROHS COMPLIANT





## FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	S2	]			
SHAPE	PLANO	PLANO			Edmund Ontice	<b>SO</b> R
SURFACE ACCURACY P-V AT 632.8nm	0.10 WAVE	0.10 WAVE			Edmund Option	<b>55</b> °
SURFACE QUALITY	10-5	10-5			Ø25.4mm HS 266T/532, 1064R HARMC	DNIC
CLEAR APERTURE	Ø22.86	Ø22.86	THIRD ANGLE PROJECTION	TITLE	BEAMSPLITTER	JI VIC
COATING APERTURE	Ø22.86	Ø22.86			52,	CLIEFT
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN mm	DWG NO	11548	SHEET 1 OF 1