NOTES:

- 1. SUBSTRATE: N-BK7
- 2. S2 TO BE PARALLEL WITH S1 TO <1 ARCMIN
- 3. COATING (APPLY ACROSS CLEAR APERTURE)


S1: VIS-NIR (400-1000nm)


S2: VIS-NIR (400-1000nm)

Rabs ≤0.25% @ 880nm Ravg ≤1.25% @ 400 - 870 nm Ravg ≤1.25% @ 890 - 1000nm



5. ROHS COMPLIANT





FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2		
SHAPE	PLANO	PLANO		
SURFACE QUALITY	60-40	60-40		
CLEAR APERTURE	22.50 x 22.50	22.50 x 22.50		
BEVEL	PROTECTED AS NEEDED	PROTECTED AS NEEDED		

		Edmund Optics®			
THIRD ANGLE PROJECTION	25mm Sq., 1mm Thick, VIS-NIR Coated λ/4 N- BK7 Window		λ/4 N-		
ALL DIMS IN	mm	DWG NO	23414	SHEET 18 OF 48	